

## IrDA adaptor compatibility with Microsoft Windows

### Hardware: OdaLog L2 / RTx

#### Issue:

In 2015, when Windows upgraded their software to Windows 10 it did not include drivers for IrDA capability and this feature was no longer operational in a large range of instruments. It was necessary to remain in an earlier version of Windows for IrDA functionality until an upgrade became available.

#### Solution:

We are pleased to announce that the *Anniversary update* of Windows 10 (version 1607; code name: RedStone 1; released 18<sup>th</sup> July 2016) as again returned compatibility with both IrDA and IrDC adaptors.

Thus OdaStat software works successfully to make a connection to OdaLog unit.

Please advise any of your customers who have been experiencing problems with older version of Windows 10 to upgrade to version 1607.

Also, any customers who have been holding off the upgrade can now safely progress to version 1607 with safety.

We will follow up this compatibility information with recent *Creator update* of Windows 10 (version 1703; code name: RedStone 2; released 11<sup>th</sup> April 2017).

#### Related Links:

Click on <https://www.microsoft.com/en-au/windows> for instructions on how to download Windows 10 or update from an earlier version.